



Calibration Laboratory Cert: 5518.01

ISO/IEC 17025:2017 and ANSI/NCSL Z540.1-1994 Accredited Calibration Certificate

Customer Address: Rental

Order:

Certificate: A24101403DR

Product: Power Sensor
 Manufacturer: Agilent
 Model: E9327A
 Serial: 1498693

Notes: Frequency Range 50MHz to 18GHz
 Tested Fully Functional to 20GHz

Date of Report: 10/14/2024
 Date of Calibration: 10/14/2024
 Next Calibration:

The next calibration date is defined by the equipment user/owner. We recommend recalibration annually.

The above instrument was tested and found to be within the manufacturer's specification. The results of the tests performed are held on file at The EMC Shop. The calibration was carried out in accordance with the general requirements of ISO/IEC 17025-2017, IEC 61000-4-6 using laboratory standards which are traceable to the SI International System of Quantities through the National Institute of Standards and Technology (NIST), and or other Accredited bodies except where none exist. Tests are carried out in environmental conditions controlled to the extent appropriate to the instrument's specification. This certificate shall not be reproduced except in full without the written approval of the laboratory. The uncertainty results meet the requirements of the ISO/IEC 17025-2017 standard and ILAC Doc.P14.

Ambient Conditions of Laboratory

Temperature (°C): **21**
 Relative Humidity (%): **37**

Technician: **Dan Raines**

Technician Signature: _____



Calibration Equipment				
Model	Description	Serial Number	Certificate #	Due Date
E5071C	Network Analyzer	MY46214012	A24041501DR	4/15/2026
85052D	Cal Kit	MY43253324	A24041502DR	4/15/2026
E4416B	Power Meter	MY56220009	A23042001DR	4/20/2025
N5183A	Signal Generator	MY50140101	753924	3/22/2027

Calibration method used: IEC 61000-4-6

VSWR measurements made are not accredited

Condition as found:	IN tolerance
Condition as left:	IN tolerance

Measuring Uncertainties	
Insertion Loss	± 0.2 dB
RF Power Level	± 0.3 dB
Power Amplifier Gain Linearity	± 0.44 dB

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%



1dB step Accuracy tested at 50MHz

Nominal (dBm)	Actual (dBm)	Difference (dBm)	Tol. Error (dBm)	Verdict
0.00	0.20	0.20	±1.00	PASS
-1.00	-0.80	0.20	±1.00	PASS
-2.00	-1.80	0.20	±1.00	PASS
-3.00	-2.80	0.20	±1.00	PASS
-4.00	-3.80	0.20	±1.00	PASS
-5.00	-4.80	0.20	±1.00	PASS
-6.00	-5.80	0.20	±1.00	PASS
-7.00	-6.80	0.20	±1.00	PASS
-8.00	-7.80	0.20	±1.00	PASS
-9.00	-8.80	0.20	±1.00	PASS
-10.00	-9.80	0.20	±1.00	PASS

1dB step Accuracy tested at 20 GHz

Nominal (dBm)	Actual (dBm)	Difference (dBm)	Tol. Error (dBm)	Verdict
0.00	-0.20	-0.20	±1.00	PASS
-1.00	-1.20	-0.20	±1.00	PASS
-2.00	-2.20	-0.20	±1.00	PASS
-3.00	-3.20	-0.20	±1.00	PASS
-4.00	-4.20	-0.20	±1.00	PASS
-5.00	-5.20	-0.20	±1.00	PASS
-6.00	-6.20	-0.20	±1.00	PASS
-7.00	-7.20	-0.20	±1.00	PASS
-8.00	-8.20	-0.20	±1.00	PASS
-9.00	-9.20	-0.20	±1.00	PASS
-10.00	-10.30	-0.30	±1.00	PASS



Power Level Accuracy versus Frequency tested at 14dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	14.00	±1.00	PASS
100 MHz	14.00	±1.00	PASS
300 MHz	13.90	±1.00	PASS
1 GHz	14.00	±1.00	PASS
3 GHz	14.00	±1.00	PASS
6 GHz	14.00	±1.00	PASS
10 GHz	14.00	±1.00	PASS
15 GHz	13.90	±1.00	PASS
18 GHz	13.80	±1.00	PASS
20 GHz	13.70	±1.00	PASS

Power Level Accuracy versus Frequency tested at 10dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	10.10	±1.00	PASS
100 MHz	10.10	±1.00	PASS
300 MHz	10.10	±1.00	PASS
1 GHz	10.20	±1.00	PASS
3 GHz	10.10	±1.00	PASS
6 GHz	10.00	±1.00	PASS
10 GHz	10.00	±1.00	PASS
15 GHz	9.90	±1.00	PASS
18 GHz	9.80	±1.00	PASS
20 GHz	9.80	±1.00	PASS



Power Level Accuracy versus Frequency tested at 0dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	0.20	±1.00	PASS
100 MHz	0.20	±1.00	PASS
300 MHz	0.10	±1.00	PASS
1 GHz	0.20	±1.00	PASS
3 GHz	0.00	±1.00	PASS
6 GHz	0.00	±1.00	PASS
10 GHz	-0.10	±1.00	PASS
15 GHz	-0.20	±1.00	PASS
18 GHz	-0.30	±1.00	PASS
20 GHz	-0.20	±1.00	PASS

Power Level Accuracy versus Frequency tested at -10dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	-9.80	±1.00	PASS
100 MHz	-9.80	±1.00	PASS
300 MHz	-9.90	±1.00	PASS
1 GHz	-9.90	±1.00	PASS
3 GHz	-10.00	±1.00	PASS
6 GHz	-10.10	±1.00	PASS
10 GHz	-10.20	±1.00	PASS
15 GHz	-10.30	±1.00	PASS
18 GHz	-10.40	±1.00	PASS
20 GHz	-10.30	±1.00	PASS



Power Level Accuracy versus Frequency tested at -20dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	-19.80	±1.00	PASS
100 MHz	-19.80	±1.00	PASS
300 MHz	-19.80	±1.00	PASS
1 GHz	-19.80	±1.00	PASS
3 GHz	-20.00	±1.00	PASS
6 GHz	-20.10	±1.00	PASS
10 GHz	-20.10	±1.00	PASS
15 GHz	-20.30	±1.00	PASS
18 GHz	-20.30	±1.00	PASS
20 GHz	-20.30	±1.00	PASS

Power Level Accuracy versus Frequency tested at -30dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	-29.70	±1.00	PASS
100 MHz	-29.70	±1.00	PASS
300 MHz	-29.70	±1.00	PASS
1 GHz	-29.80	±1.00	PASS
3 GHz	-29.90	±1.00	PASS
6 GHz	-30.00	±1.00	PASS
10 GHz	-30.00	±1.00	PASS
15 GHz	-30.20	±1.00	PASS
18 GHz	-30.20	±1.00	PASS
20 GHz	-30.20	±1.00	PASS



Power Level Accuracy versus Frequency tested at -40dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	-39.30	±1.00	PASS
100 MHz	-39.40	±1.00	PASS
300 MHz	-39.40	±1.00	PASS
1 GHz	-39.50	±1.00	PASS
3 GHz	-39.60	±1.00	PASS
6 GHz	-39.60	±1.00	PASS
10 GHz	-39.60	±1.00	PASS
15 GHz	-39.50	±1.00	PASS
18 GHz	-39.50	±1.00	PASS
20 GHz	-39.40	±1.00	PASS

Power Level Accuracy versus Frequency tested at -50dBm

Frequency	Actual (dBm)	Tol. Error (dBm)	Verdict
50 MHz	-49.10	±1.00	PASS
100 MHz	-49.10	±1.00	PASS
300 MHz	-49.10	±1.00	PASS
1 GHz	-49.20	±1.00	PASS
3 GHz	-49.20	±1.00	PASS
6 GHz	-49.20	±1.00	PASS
10 GHz	-49.20	±1.00	PASS
15 GHz	-49.20	±1.00	PASS
18 GHz	-49.20	±1.00	PASS
20 GHz	-49.10	±1.00	PASS



VSWR

