



Calibration Laboratory Cert: 5518.01

ISO/IEC 17025:2017 and ANSI/NCSL Z540.1-1994 Accredited Calibration Certificate

Customer Address:	Rental
Order / RMA:	
Certificate:	A24092001DR
Product:	Amplifier
Manufacturer:	The EMC Shop
Model:	5155F
Serial:	1002
Notes:	Frequency Range: 800MHz to 2.5GHz
	CSV files available upon request
Date of Report:	9/20/2024
Date of Calibration:	9/20/2024
Nevt Calibration	

The next calibration date is defined by the equipment user/owner.

The above instrument was tested and found to be within the Manufacturer's specifications at the tested parameters. The results of the tests performed are held on file at The EMC Shop. The calibration was carried out in accordance with the general requirements of ISO/IEC 17025-2017 and ANSI Z-540-1 using laboratory standards which are traceable to the SI International System of Quantities through the National Institute of Standards and Technology (NIST), and or other Accredited bodies except where none exist. Tests are carried out in environmental conditions controlled to the extent appropriate to the instrument's specification. This certificate shall not be reproduced except in full without the written approval of the laboratory. The uncertainty results meet the requirements of the ISO/IEC 17025-2017 standard and ILAC Doc.P14. Statements of conformity (e.g. Pass or Fail) are made in accordance with Simple Acceptance decision rules as defined in ILAC G8 with a TUR of 4:1 or greater. The customer is responsible for considering whether the inclusion of the uncertainties shown on the certificate would prevent their use of the equipment based on their risk evaluations. Results are accredited unless annotated with an asterisk "*". The results presented are only applicable to the Model/Serial number shown above.

Ambient Conditions of Labo	ratory	Technician:	Dan Raines
Temperature (°C):	21		
Relative Humidity (%):	36	Technician Signature:	



	Calibration	Equipment		
Model	Description	Serial Number	Certificate #	Due Date
ZNB 8	Rohde&Schwarz Vect. Netw. Ana.	102017	S220805-021	8/9/2025
ZV-Z21	Calibration Kit (50Ω)	100800	0001A300706445	7/27/2025

Calibration method used: IEC 61000-4-3		Calibration method used: IEC 61000-4-3	
--	--	--	--


Condition as found:	IN tolerance
Condition as left:	IN tolerance

Measuring Uncertainties		
Insertion Loss	0.44 dB	


The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%



Gain Through Band




Input VSWR Tolerance: ≤2.0:1





P1dB 800MHz





P1dB 2.5GHz



^{*}CSV Files for Characterizations are available upon request.

Frequency	Gain Through Band	VSWR at Input
(Hz)	(dBm)	(U)
800,000,000	65.95	1.34
808,500,000	66.95	1.34
817,000,000	67.65	1.34
825,500,000	68.26	1.33
834,000,000	68.71	1.33
842,500,000	68.90	1.32
851,000,000	68.89	1.31
859,500,000	68.73	1.30
868,000,000	68.58	1.29
876,500,000	68.54	1.28
885,000,000	68.56	1.26
893,500,000	68.59	1.25
902,000,000	68.74	1.24
910,500,000	68.83	1.22
919,000,000	68.65	1.21
927,500,000	68.28	1.19
936,000,000	67.88	1.18
944,500,000	67.27	1.16
953,000,000	66.56	1.14
961,500,000	66.00	1.13
970,000,000	65.36	1.12
978,500,000	64.65	1.11
987,000,000	64.02	1.10
995,500,000	63.30	1.10
1,004,000,000	62.69	1.10
1,012,500,000	62.31	1.10
1,021,000,000	62.07	1.11
1,029,500,000	62.13	1.12
1,038,000,000	62.14	1.13
1,046,500,000	62.43	1.15
1,055,000,000	62.46	1.16
1,063,500,000	62.58	1.18
1,072,000,000	62.61	1.20
1,080,500,000	62.67	1.21
1,089,000,000	62.50	1.22
1,097,500,000	62.33	1.24
1,106,000,000	62.10	1.25
1,114,500,000	61.74	1.26
1,123,000,000	61.59	1.27
1,131,500,000	61.34	1.28

1,140,000,000	61.16	1.29
1,148,500,000	61.03	1.30
1,157,000,000	60.82	1.30
1,165,500,000	60.50	1.31
1,174,000,000	60.32	1.31
1,182,500,000	60.16	1.31
1,191,000,000	60.00	1.31
1,199,500,000	59.92	1.31
1,208,000,000	60.07	1.30
1,216,500,000	60.17	1.30
1,225,000,000	60.52	1.29
1,233,500,000	60.83	1.29
1,242,000,000	61.27	1.28
1,250,500,000	61.66	1.27
1,259,000,000	62.02	1.27
1,267,500,000	62.42	1.25
1,276,000,000	62.87	1.24
1,284,500,000	63.40	1.24
1,293,000,000	63.86	1.23
1,301,500,000	64.53	1.22
1,310,000,000	64.97	1.21
1,318,500,000	65.29	1.20
1,327,000,000	65.57	1.19
1,335,500,000	65.72	1.18
1,344,000,000	65.97	1.18
1,352,500,000	66.16	1.17
1,361,000,000	66.62	1.17
1,369,500,000	66.96	1.16
1,378,000,000	67.53	1.16
1,386,500,000	68.11	1.15
1,395,000,000	68.56	1.14
1,403,500,000	69.04	1.14
1,412,000,000	69.53	1.14
1,420,500,000	69.74	1.13
1,429,000,000	70.08	1.13
1,437,500,000	70.33	1.13
1,446,000,000	70.59	1.13
1,454,500,000	70.76	1.13
1,463,000,000	70.72	1.13
1,471,500,000	70.74	1.13
1,480,000,000	70.73	1.13
1,488,500,000	70.51	1.13
1,497,000,000	70.25	1.13
1,505,500,000	69.91	1.13

1,514,000,000	69.66	1.13
1,522,500,000	69.20	1.13
1,531,000,000	68.81	1.13
1,539,500,000	68.45	1.13
1,548,000,000	68.11	1.13
1,556,500,000	67.66	1.13
1,565,000,000	67.24	1.13
1,573,500,000	66.67	1.12
1,582,000,000	66.07	1.12
1,590,500,000	65.45	1.12
1,599,000,000	65.00	1.12
1,607,500,000	64.83	1.11
1,616,000,000	64.64	1.11
1,624,500,000	64.50	1.10
1,633,000,000	64.25	1.10
1,641,500,000	64.08	1.10
1,650,000,000	63.75	1.10
1,658,500,000	63.53	1.10
1,667,000,000	63.32	1.10
1,675,500,000	63.17	1.10
1,684,000,000	63.02	1.10
1,692,500,000	63.07	1.11
1,701,000,000	62.99	1.11
1,709,500,000	62.59	1.12
1,718,000,000	62.36	1.12
1,726,500,000	62.10	1.13
1,735,000,000	61.76	1.13
1,743,500,000	61.32	1.14
1,752,000,000	60.91	1.15
1,760,500,000	60.26	1.15
1,769,000,000	59.87	1.16
1,777,500,000	59.68	1.17
1,786,000,000	59.53	1.17
1,794,500,000	59.67	1.18
1,803,000,000	60.18	1.18
1,811,500,000	60.52	1.19
1,820,000,000	60.65	1.19
1,828,500,000	60.54	1.20
1,837,000,000	60.45	1.20
1,845,500,000	60.37	1.21
1,854,000,000	60.56	1.21
1,862,500,000	60.89	1.21
1,871,000,000	61.30	1.21
1,879,500,000	61.48	1.21

1,888,000,000 61.57	
1,000,000,000	1.22
1,896,500,000 61.59	1.22
1,905,000,000 61.59	1.22
1,913,500,000 61.58	1.22
1,922,000,000 61.43	1.22
1,930,500,000 61.32	1.22
1,939,000,000 61.18	1.22
1,947,500,000 61.05	1.22
1,956,000,000 60.97	1.22
1,964,500,000 60.91	1.21
1,973,000,000 60.84	1.21
1,981,500,000 60.75	1.21
1,990,000,000 60.50	1.21
1,998,500,000 60.24	1.21
2,007,000,000 60.06	1.20
2,015,500,000 59.83	1.20
2,024,000,000 59.55	1.20
2,032,500,000 59.49	1.19
2,041,000,000 59.51	1.19
2,049,500,000 59.60	1.19
2,058,000,000 59.53	1.19
2,066,500,000 59.48	1.18
2,075,000,000 59.50	1.18
2,083,500,000 59.53	1.17
2,092,000,000 59.46	1.17
2,100,500,000 59.61	1.16
2,109,000,000 59.74	1.16
2,117,500,000 59.72	1.16
2,126,000,000 59.67	1.15
2,134,500,000 59.43	1.14
2,143,000,000 59.28	1.14
2,151,500,000 59.19	1.13
2,160,000,000 59.10	1.13
2,168,500,000 59.11	1.12
2,177,000,000 59.02	1.11
2,185,500,000 59.10	1.11
2,194,000,000 59.25	1.10
2,202,500,000 59.37	1.09
2,211,000,000 59.73	1.09
2,219,500,000 60.26	1.08
2,228,000,000 60.67	1.07
2,236,500,000 60.88	1.06
2,245,000,000 61.08	1.05
2,253,500,000 61.47	1.05

2,262,000,000	61.66	1.04
2,270,500,000	61.96	1.03
2,279,000,000	62.02	1.03
2,287,500,000	62.26	1.03
2,296,000,000	62.31	1.03
2,304,500,000	62.40	1.04
2,313,000,000	62.21	1.04
2,321,500,000	62.06	1.05
2,330,000,000	62.01	1.06
2,338,500,000	62.08	1.07
2,347,000,000	62.04	1.08
2,355,500,000	62.15	1.09
2,364,000,000	62.29	1.10
2,372,500,000	62.31	1.11
2,381,000,000	62.12	1.12
2,389,500,000	62.08	1.13
2,398,000,000	61.84	1.14
2,406,500,000	61.91	1.15
2,415,000,000	61.91	1.16
2,423,500,000	62.13	1.16
2,432,000,000	62.21	1.17
2,440,500,000	62.12	1.18
2,449,000,000	61.97	1.19
2,457,500,000	61.79	1.20
2,466,000,000	61.77	1.21
2,474,500,000	61.54	1.22
2,483,000,000	61.20	1.22
2,491,500,000	61.10	1.23
2,500,000,000	61.04	1.24

End of Calibration Report: Amplifier cert. rev. 1